



InertSustainSwift™
GL Sciences

Introduction for a new HPLC column

InertSustainSwift™ C18

Evolutional high performance for high throughput analysis . . .


InertSustainSwift C18 is an optimal column for high throughput analysis designed for elution time to be faster and shorten an equilibration time also in gradient analysis. Moreover, it has the inactive processing technology acquired by InertSustain C18, and has become a column which combines high durability and low bleeding.



Physical Properties

InertSustainSwift™ C18

- Silica : Newly Developed ES Silica Gel
- Particle Size : 1.9 μm , 3 μm , 5 μm
- Surface Area : 200 m^2/g
- Pore Size : 200 $\text{\AA}$ (20 nm)
- Pore Volume : 1.00 mL/g
- Bonded Phase : Octadecyl Groups
- End-capping : Complete
- Carbon Loading : 9.0 %
- USP Code : L1
- pH Range : 1.0 ~ 10.0



Evolved Surface Silica

Features

↑ Wide pH range (pH1 to10)

Evolved Surface Silica

↑ Shorten an equilibration time

Optimized Surface Area (200 m²/g) and Pore size (200 Å)

↑ Suitable retentivity

Perfect Controlled Carbon Loading (9.0 %)

↑ High inertness and Lower bleeding

Latest Production Process


↑ Lower pressure

Traditional GL Sciences' Silica Production Technology

↑ Appropriate performance for LC/MS/MS


High theoretical plate number and Lower bleeding

Superior Features



InertSustainSwift C18 (3 µm)


Shorten a retention time !



InertSustain C18 (3 µm)

Conditions

Eluent : A) CH₃OH
B) H₂O
A/B = 80/20, v/v
Column Size : 3 µm, 150 × 2.1mm I.D.
Flow Rate : 0.3 mL/min
Col. Temp. : 40 °C
Detection : UV 254 nm
Sample : 1. Uracil
2. Toluene
3. Ethylbenzene
4. Propylbenzene
5. *n*-Butylbenzene
6. *n*-Amylbenzene




InertSustainSwift C18 (1.9 µm)

High theoretical plate number

Lower column pressure

**N1:6500
N2:9900
27.2 MPa**



Acquity BEH (1.7 µm)

**N1:1300
N2:9100
35.7 MPa**

Conditions

Eluent : A) CH₃OH
B) H₂O
A/B = 30/70, v/v
Column Size : 50 × 2.1 mm I.D.
Flow Rate : 0.4 mL/min
Col. Temp. : 40 °C
Detection : UV 254 nm
Sample : 1. Pyridine
2. Phenol

Important contents for High Throughput Analysis

Rapid equilibration

Suitably retentivity


High inertness and Lower bleeding

Incomplete Equilibration

If equilibration time is not enough in gradient analysis . . .

(Possible symptoms)



Bad reproducibility in high water solubility components especially




Wave of base line



Unstable sensitivity in LC/MS/MS



Comparison Test for Equilibration time (Conditions)



Conditions

Column Size : 3 μ m, 150 $\times$ 2.1 mm I.D.

Eluent : A) H₂O B) CH₃CN

A/B = 95/5 (8 min hold) – 5 min - 0/100 –

5 min- 0/100 - 0.1min - 95/5 – **Equilibration time** , v / v


Flow Rate : 0.2 mL/min

Col. Temp. : 40 °C


Detection : UV 215 nm

Sample : Water-soluble components (Pesticides)


1. Methamidophos



2. Acephate






Comparison Testing for Equilibration time (Results)



InertSustainSwift C18 has faster stabilization !

100% Water Condition

InertSustainSwift C18 can meet with 100% water condition easily.



Variation of retention time may be occurred by that ODS column is not equilibrated with high polar solvents easily such as water in gradient analysis.

However . . .

InertSustainSwift C18 can be used for 100% Water condition and equilibrated faster in gradient analysis by optimal carbon loading and pore size.

Testing Procedure:


- 1) 100 % water is introduced into column over 20 minutes.
- 2) Conduct Analysis
- 3) Stop flow for 15 minutes.
- 4) 100 % water is introduced again into column over 30 minutes.
- 5) Stop flow for 15 minutes again.
- 6) Conduct Analysis

Conditions


Column : InertSustainSwift C18
(5 μ m, 4.6 $\times$ 250)
Eluent : 100 % H₂O
Flow rate : 1.0 mL/min
Col. Temp. : 40 $^{\circ}$ C
Detection : UV 254 nm
Sample : 1.Cytosine 2.Uracil
3.Guanine 4.Thymine
5.Adanine

Suitable Retentivity


● Comparison data InertSustainSwift C18 with InertSustain C18 and C8



InertSustainSwift C18



InertSustain C8



InertSustain C18

Conditions

Column Size

: 3 μ m, 150 $\times$ 2.1 mm I.D.

Eluent

: CH₃OH/H₂O = 80/20, v/v

Flow Rate

: 0.3 mL/min

Col. Temp.

: 40 $^{\circ}$ C

Detection

: UV 254 nm

Sample

: 1: Uracil

2: Toluene

3: Ethylbenzene

4: Propylbenzene


5: *n*-Butylbenzene

6: *n*-Amylbenzene


Suitable Retentivity

- Hydrophobic contaminations also are eluted faster !

High
hydrophobic
ODS column




InertSustainSwift C18



Conditions

Column Size : 3 μ m, 150 $\times$ 2.1mm I.D.
Eluent : A) CH₃CN
 B) 0.1 % HCOOH, 50 mM HCOONH₄ in H₂O
 A/B = 65/35, v/v
Flow Rate : 0.3 mL/min
Col. Temp. : 40 °C
Detection : UV 225 nm
Injection Vol : 5 μ L
Sample : 1. Ethylbenzene

Adsorption of Analytes



After disconnect the column joints, confirm to the packing of inlet !

Conditions

Eluent	: CH ₃ CN
Flow rate	: 1 mL/min
Col. Temp	: Room temperature
Injection Vol.	: 100 μ L
Sample	: Coffee sample


Before injecting

After injecting

After washing
by CH₃CN for 10min

Existing C18 column



InertSustainSwift C18



Clean

Highly Inertness

Column size: 50 mm × 2.1 mm I.D.




Conditions


Eluent : A) CH₃CN
B) 25mM K₂HPO₄ (pH 7.0, KH₂PO₄)
A/B = 30/70,v/v
Flow Rate : 0.4 mL/min
Col. Temp. : 40 °C
Detection : UV 230 nm
Injection Vol : 0.5 μ L
Sample : 1:Uracil
2:Pyridine
3:Phenol
4:Berberine chloride
5:Dextromethorphan

Highly Inertness

InertSustainSwift C18



Zorbax Eclipse Plus C18



Conditions

Eluent : A) CH₃CN
 B) 25mM K₂HPO₄ (pH 7.0, KH₂PO₄)
 A/B = 60/40, v/v
 Flow Rate : ※1
 Col. Temp. : 40 °C
 Detection : UV 220 nm
 Injection Vol : ※2

Sub 2 μm (2.1 × 50)
 ※1 : 0.4 mL/min ※2 : 0.5 μL
 3 μm (2.1 × 150)
 ※1 : 0.3 mL/min ※2 : 1 μL
 5 μm (4.6 × 250)
 ※1 : 1.0 mL/min ※2 : 2 μL

Sample:


1. Chlorpheniramine
2. Triprolidine
3. Homochlorcyclizine
4. Hydroxyzine
5. Clemastine

Reliable Reproducibility

Column Size : InertSustainSwift C18 3 μm , 150 $\times$ 4.6 mm I.D.


Strong Basic Compound Test

1. Dextromethorphan




Strong Acidic Compound Test

1. Brilliant Blue FCF



Strong Chelating Compound Test

1. Hinokitiol



Ion Suppression in LC/MS(/MS) analysis



What is Ion Suppression ?

Ionized mass number at the same time has the limit in LC/MS(/MS) analysis.

If foreign substances were included in sample and the retention time was same with a target component, the sensitivity for target component will be decreased due to foreign substances.


⊕ : Foreign substances

⊕ : Target component




Good sensitivity for target component

Preventing Ion Suppression




Offer Good Peak Shapes (1)

InertSustainSwift C18 can offer good peak shapes without adsorption due to high inertness. And also, high theoretical plate number will be achieved at low column back pressure due to 1.9 μm particle size.



InertSustainSwift C18 (1.9 μm)



Acquity BEH (1.7 μm)


Conditions

Eluent	: A) CH ₃ OH B) H ₂ O A/B = 30/70, v/v
Column Size	: 50 × 2.1 mm I.D.
Flow Rate	: 0.4 mL/min
Col. Temp.	: 40 °C
Detection	: UV 254 nm
Sample	: 1. Pyridine 2. Phenol

Offer Good Peak Shapes (2)

Basic compound adsorbed with remained silanol can be analyzed easily without any tailing due to highly inertness.

● Application for basic drug under formic acid mobile phase condition




Conditions


System : Nexera
LCMS-8030 plus
Column : InertSustainSwift C18
(3 μ m, 150 $\times$ 2.1 mm I.D.)
Eluent : A) 0.1% HCOOH in H₂O
B) 0.1% HCOOH in CH₃CN
A/B = 95/5 - 5 min - 0/100
- 5min - 0/100 - 0.1min
- 95/5 - 5min - 95/5 -,v/v
Flow Rate : 0.2 mL/min
Col. Temp. : 40 $^{\circ}$ C
Detection : LC/MS (ESI)

Confirmation Testing for Influence of Buffer (Conditions)

Concentration of ammonium acetate was changed in stages and then, peak shapes of each stage was confirmed if adsorption was occurred.



Dibucaine (pKa 8.5)



Promethazine (pKa 9.1)

Conditions

System : Nexera
LCMS-8030 plus
Column size : (3 μ m, 150 $\times$ 2.1 mm I.D.)
Eluent : A) Ammonium Acetate in H₂O
B) Ammonium Acetate in CH₃OH
A/B = 30/70 ,v/v
Flow Rate : 0.2 mL/min
Col. Temp. : 40 $^{\circ}$ C
Detection : LC/MS/MS (ESI, Positive, MRM)

Sample: each 0.1 (mg / mL)


Q1 > Q3

- | | |
|-----------------|-----------------------|
| 1. Dibcaine | : 345.20 > 272.15 (+) |
| 2. Promethazine | : 287.05 > 86.15 (+) |


Confirmation Test for Influence of Buffer (Results)

InertSustainSwift C18

1mM Ammonium acetate



5mM Ammonium acetate




1. Dibucaine (pKa 8.5)
2. Promethazine (pKa 9.1)

(Results)


Good result !
Peak shapes were not changed

CAPCELLPAK MGIII C18

1mM Ammonium acetate



5mM Ammonium acetate



Peak shapes were improved from 1mM to 5mM, however the sensitivity were low due to adsorption.

(Possible symptoms)


Incorrect calibration curve
Decreasing sensitivity

Extremely Low Bleeding (1)

It is known that an unreacted reagents and an product of the reagents from chemical modification in the production process of ODS provide a column bleeding. These causes of the column bleeding are possible to be eluted in the gradient analysis and make a ion suppression.

However, InertSustainSwift C18 offer **“High inertness”** and **“Low bleeding”**.

(x1,000,000,000) TIC(+@1




Conditions

System : Nexera
LCMS-8030 plus
Column Size : 3 μ m, 150 $\times$ 2.1 mm I.D.)
Eluent : A) 0.1 % HCOOH in H₂O
B) 0.1 % HCOOH in CH₃CN
A/B = 95/5 - 5 min - 0/100 - 5 min - 0/100 -
0.1 min - 95/5 - 5 min - 95/5, v/v
Flow Rate : 0.2 mL/min
Col. Temp. : 40 °C
Detection : LC/MS (ESI, Positive,
Mass Range: 50-1000)
Q1 SCAN : m/z 50 - 1000

Extremely Low Bleeding (2)


Existing column



Conditions

System : Nexera
 LCMS-8030 plus
 Column Size : 3 μ m, 150 $\times$ 2.1 mm I.D.)
 Eluent : A) 0.1 % HCOOH in H₂O
 B) 0.1 % HCOOH in CH₃CN
 A/B = 95/5 - 5 min - 0/100 - 5 min - 0/100

InertSustainSwift C18




Flow Rate : 0.2 mL/min
 Col. Temp. : 40 °C
 Detection : LC/MS (ESI, Positive,
 Mass Range: 50-1000)
 Q1 SCAN : m/z 50 - 1000

Application (1)

● Analysis of β -adrenergic blocking agent (β -blocker)

Conditions

System : Nexera
LCMS-8030 plus
Column : 1.9 μ m, 2.1 x 50 mm
Eluent : A) 10 mM Ammonium Formate in H₂O
B) 10 mM Ammonium Formate in CH₃OH
A/B = 70/30 – 0.3 min - 40/60 – 0.5min - 0/100 – 0.1min – 0/100 – 0.01min -70/30 – 0.5min - 70/30 ,v/v
Flow Rate : 0.6 mL/min
Col. Temp. : 40 °C
Detection : LC/MS/MS (ESI, Positive, Negative MRM)



Sample: each 100 (μ g/L)

Q1 > Q3

1. Acebutolol	: 337.10 > 116.05 (+)
2. Atenolol	: 267.25 > 145.00 (+)
3. Labetalol	: 329.00 > 161.95 (+)
4. Nadolol	: 310.05 > 254.00 (+)
5. Pindolol	: 249.80 > 116.00 (+)


Comparison Test Contents

- 1) Comparison of Retentivity**
- 2) Comparison of Separation/Selectivity**
- 3) Comparison of InertSustainSwift C18 with the other brand**
 - a) Hydrophobicity Test**
 - b) Chelating Compounds Test**
 - c) Weak Acidic/Basic Compounds Test**
 - d) Acidic Compounds Test**
 - e) Basic Compounds Test**
- 4) Test for Influence of Buffer**
- 5) Durability Test**


1) Comparison of Retentivity

Column dimensions : 3 μ m 150 $\times$ 2.1 mm I.D.

Sample : Alkyl benzenes




2) Comparison of Separation/Selectivity



3) Comparison of InertSustainSwift C18 with the other brand

(Test conditions)

a) Hydrophobicity Test






R: 2. Toluene

3. Ethylbenzene

4. Propylbenzene

5. Butylbenzene

6. Amylbenzene



Conditions

Eluent : CH₃OH/H₂O = 80/20


Flow Rate : 5μm (4.6 x 250mm) 1.0mL/min, Inj. Vol. 5μL
 3μm (2.1 x 150mm) 0.3mL/min, Inj. Vol. 1μL
 1.9μm (2.1 x 50mm) 0.4mL/min, Inj. Vol. 0.5μL

Col. Temp. : 40 °C

Detection : UV 254 nm

Sample : 1: Uracil 2: Toluene 3: Ethylbenzene
 4: Propylbenzene 5: *n*-Butylbenzene
 6: *n*-Amylbenzene

b) Chelating Compounds Test



Hinokitiol

Conditions

Eluent : A) CH₃CN
 B) 0.1% H₃PO₄
 A/B = 40/60

Flow Rate : 5μm (4.6 x 250mm) 1.0mL/min, Inj. vol. 3μL
 3μm (2.1 x 150mm) 0.3mL/min, Inj. vol. 1μL
 1.9μm (2.1 x 50mm) 0.4mL/min, Inj. vol. 0.5μL

Col. Temp. : 40 °C


Detection : UV 254 nm

Sample : 1: β-Thujaplicin (Hinokitiol)


3) Comparison of InertSustainSwift C18 with the other brand

(Test conditions)

c) Weak Acidic/Basic Compounds Test



Pyridine




Phenol

Conditions

Eluent	: A) CH ₃ OH B) H ₂ O A/B = 30/70
Flow Rate	: 5um (4.6 x 250mm) 1.0mL/min, Inj. vol. 1uL 3um (2.1 x 150mm) 0.3mL/min, Inj. vol. 1uL 1.9um (2.1 x 50mm) 0.3mL/min, Inj. vol. 0.5uL
Col. Temp.	: 40 °C
Detection	: UV 254 nm
Sample	: 1: Uracil 2: Pyridine 3: Phenol

d) Acidic Compounds Test



Brilliant Blue FCF


Conditions

Eluent	: A) CH ₃ CN B) 0.1% H ₃ PO ₄ A/B = 25/75, v/v
Flow Rate	: 5um (4.6 x 250mm) 1.0mL/min, Inj. vol. 3uL 3um (2.1 x 150mm) 0.3mL/min, Inj. vol. 1uL 1.9um (2.1 x 50mm) 0.4mL/min, Inj. vol. 0.5uL
Col. Temp.	: 40 °C
Detection	: UV 254 nm
Sample	: 1: Brilliant Blue FCF 2: 4-Ethylbenzoic acid 3: Salicylic acid


3) Comparison of InertSustainSwift C18 with the other brand

(Testing conditions)

e) Basic Compounds Test (1)



Berberine chloride




Dextromethorphan


Conditions

Eluent	: A) CH ₃ CN B) 25 mM K ₂ HPO ₄ (pH 7.0, KH ₂ PO ₄) A/B = 30/70, v/v
Flow Rate	: 5μm (4.6 x 250mm) 1.0mL/min, Inj. vol. 2μL 3μm (2.1 x 150mm) 0.3mL/min, Inj. vol. 1μL 1.9μm (2.1 x 50mm) 0.4mL/min, Inj. vol. 0.5μL
Col. Temp.	: 40 °C
Detection	: UV 230 nm
Sample	: 1:Uracil 2:Pyridine 3:Phenol 4:Berberine chloride 5:Dextromethorphan


e) Basic Compounds Test (2)




Chlorpheniramine




Homochlorcyclizine



Clemastine



Triprolidine



Hydroxyzine

Conditions

Eluent	: A) CH ₃ CN B) 25mM K ₂ HPO ₄ (pH 7.0, KH ₂ PO ₄) A/B = 60/40, v/v
Flow Rate	: 5μm (4.6 x 250mm) 1.0mL/min, Inj. vol. 5μL 3μm (2.1 x 150mm) 0.3mL/min, Inj. vol. 1μL 1.9μm (2.1 x 50mm) 0.4mL/min, Inj. vol. 0.5μL
Col. Temp.	: 40 °C
Detection	: UV 220 nm
Sample	: 1: Chlorpheniramine 2: Triprolidine 3: Homochlorcyclizine 4: Hydroxyzine 5: Clemastine

3) Comparison of InertSustainSwift C18 with the other brand

Comparison List (Sub 2µm)

Description	Particle size (µm)	Surface area (m ² /g)	Pore size (Å)	Pore Volume (mL/g)	Carbon Loading (%)	pH
InertSustainSwift C18	1.9	200	200	1	9	1 to 10
Zorbax Eclipse Plus C18 (Agilent)	1.8	160	95	—	8	2 to 9
Acquity BEH C18 (Waters)	1.7	185	135	0.7	18	1 to 12
Hypersil GOLD C18 (Thermo Fisher Scientific)	1.9	220	175	—	10	1 to 11
Triart C18 (YMC)	1.9	—	120	—	20	1 to 12
L-column2 ODS (CERI)	2	340	120	—	17	2 to 9
Titan C18 (Sigma-Aldrich)	1.9	410	80	0.76	13.5	2 to 8

a) Hydrophobicity

























b) Chelating

c) Weak Acidic/Basic

d) Acidic

e) Basic (1)

e) Basic (2)



a) Hydrophobicity

c) Weak Acidic/Basic


e) Basic (1)

b) Chelating


d) Acidic

e) Basic (2)


InertSustainSwift C18
























Triart C18



L-column2 ODS



Titan C18



3) Comparison of InertSustainSwift C18 with the other brand

Comparison List (Sub 2um)

Description	Inertness			Retention time	100% Water	Bleeding
	Chelating	Acidic	Basic			
InertSustainSwift C18	Great	Great	Great	1	Great	Great
Zorbax Eclipse Plus C18 (Agilent)	Poor	Poor	Normal	2	Poor	Great
Acquity BEH C18 (Waters)	Good	Normal	Normal	1	Normal	Great
Hypersil GOLD C18 (Thermo Fisher Scientific)	Poor	Normal	Poor	1	Good	Great
Triart C18 (YMC)	Good	Great	Good	2	Great	Good
L-column2 ODS (CERI)	Poor	Poor	Normal	2	Poor	Normal
Titan C18 (Sigma-Aldrich)	Poor	Great	Poor	2	Great	Great

1: Early
2: Middle
3: Late

3) Comparison of InertSustainSwift C18 with the other brand

Comparison List (3um)

Description	Particle size (µm)	Surface area (m ² /g)	Pore size (Å)	Pore Volume (mL/g)	Carbon Loading (%)	pH
InertSustainSwift C18	3	200	200	1	9	1 to 10
XBridge C18 (Waters)	3.5	185	135	0.7	18	1 to 12
Zorbax Eclipse Plus C18 (Agilent)	3.5	160	95	—	8	2 to 9
Triart C18 (YMC)	3	—	120	—	20	1 to 12
Hypersil GOLD C18 (Thermo Fisher Scientific)	3	220	175	—	10	1 to 11
Gemini NX C18 (Phenomenex)	3	375	110	1.1	14	1 to 12
Luna C18(2) (Phenomenex)	3	400	100	—	17.5	1.5 to 10
Cadenza CD-C18 (Imtakt)	3	—	120	—	—	—
CAPCELLPAK MGIII C18 (SHISEIDO)	3	300	100	—	15	2 to 10
L-column2 ODS (CERI)	3	340	120	1.1	17	2 to 9

a) Hydrophobicity

























c) Weak Acidic/Basic

e) Basic (1)

b) Chelating

d) Acidic

e) Basic (2)



a) Hydrophobicity

























c) Weak Acidic/Basic

e) Basic (1)

b) Chelating

d) Acidic

e) Basic (2)



a) Hydrophobicity

























c) Weak Acidic/Basic

e) Basic (1)

b) Chelating

d) Acidic

e) Basic (2)



3) Comparison of InertSustainSwift C18 with the other brand

Comparison List (3um)

Description	Inertness			Analytical time	100% Water	Bleeding
	Chelating	Acidic	Basic			
InertSustainSwift C18	Great	Great	Great	1	Great	Great
XBridge C18 (Waters)	Poor	Normal	Great	2	Normal	Great
Zorbax Eclipse Plus C18 (Agilent)	Poor	Normal	Poor	3	Poor	Great
Triart C18 (YMC)	Great	Normal	Great	3	Great	Good
Hypersil GOLD C18 (Thermo Fisher Scientific)	Normal	Poor	Poor	1	Good	Great
Gemini NX C18 (Phenomenex)	Normal	Poor	Good	1	Great	Great
Luna C18(2) (Phenomenex)	Poor	Poor	Poor	3	Poor	Good
Cadenza CD-C18 (Imtakt)	Normal	Good	Poor	3	Poor	Good
CAPCELLPAK MGIII C18 (SHISEIDO)	Poor	Poor	Poor	2	Great	Great
L-column2 ODS (CERI)	Normal	Normal	Normal	3	Poor	Normal


1: Early
2: Middle
3: Late

3) Comparison of InertSustainSwift C18 with the other brand


Comparison List (5um)

Description	Particle size (μm)	Surface area (m^2/g)	Pore size ($\text{\AA}$)	Pore Volume (mL/g)	Carbon Loading (%)	pH
InertSustainSwift C18	5	200	200	1	9	1 to 10
YMC-Pack Pro C18 (YMC)	5	330	120	1.05	16	2 to 8
L-column2 ODS (CERI)	5	340	120	1.1	17	2 to 9
L-column ODS (CERI)	5	340	120	1.1	17	2 to 9
Luna C18(2) (Phenomenex)	5	400	100	—	17.5	1.5 to 10
SunFire C18 (Waters)	5	340	100	1	16	2 to 8
XBridge C18 (Waters)	5	185	130	—	18	1 to 12
Hypersil BDS (ThermoFisherScientific)	5	—	130	—	11	2 to 9
Zorbax Eclipse Plus C18 (Agilent)	5	160	95	—	8	2 to 9


a) Hydrophobicity




b) Chelating




c) Weak Acidic/Basic






















d) Acidic



e) Basic (1)



e) Basic (2)



a) Hydrophobicity

























c) Weak Acidic/Basic

e) Basic (1)


b) Chelating

d) Acidic


e) Basic (2)




a) Hydrophobicity




b) Chelating




c) Weak Acidic/Basic
















d) Acidic



e) Basic (1)



e) Basic (2)



3) Comparison of InertSustainSwift C18 with the other brand


Comparison List (5um)

Description	Inertness			Analytical time	100% Water	Bleeding
	Chelating	Acidic	Basic			
InertSustainSwift C18	Great	Great	Great	1	Great	Great
YMC-Pack Pro C18 (YMC)	Normal	Poor	Poor	3	Good	-
L-column2 ODS (CERI)	Normal	Normal	Good	3	Poor	-
L-column ODS (CERI)	Normal	Poor	Normal	3	Normal	-
Luna C18(2) (Phenomenex)	Poor	Normal	Normal	2	Poor	-
SunFire C18 (Waters)	Poor	Poor	Normal	2	Poor	-
XBridge C18 (Waters)	Poor	Normal	Good	1	Poor	-
Hypersil BDS (ThermoFisherScientific)	Poor	Great	Poor	1	Poor	-
Zorbax Eclipse Plus C18 (Agilent)	Poor	Normal	Poor	2	Poor	-


1: Early
2: Middle
3: Late

4) Confirmation Test for Influence of Buffer (Conditions)

Concentration of ammonium acetate was changed in stages and then, peak shapes of each stage was confirmed if adsorption was occurred.



Dibucaine (pKa 8.5)



Promethazine (pKa 9.1)

Conditions

System : Nexera
LCMS-8030 plus
Column size : (3 μ m, 150 $\times$ 2.1 mm I.D.)
Eluent : A) Ammonium Acetate in H₂O
B) Ammonium Acetate in CH₃OH
A/B = 30/70 ,v/v
Flow Rate : 0.2 mL/min
Col. Temp. : 40 $^{\circ}$ C
Detection : LC/MS/MS (ESI, Positive, MRM)

Sample:each 0.1 (mg / mL)


Q1 > Q3
1. Dibcaine : 345.20 > 272.15 (+)
2. Promethazine : 287.05 > 86.15 (+)

Column size: 3 μ m, 2.1 $\times$ 150

Upper : Dibucaine (pka 8.5)



Under : Promethazine (pka 9.1)

4) Confirmation Test for Influence of Buffer (Results)



5) Durability Test for Acidic/Basic conditions

● Durability test for Acidic conditions





Purging Conditions

Column Size : 3 μ m, 2.1 x 150mm
 Eluent : 1% TFE (pH 1.0) or
 0.1% TFA(2.0) /
 CH₃CN : 90/10, v / v
 Col. Temp : 60°C

—●— pH 1.0
 —■— pH 2.0

● Durability test for basic conditions



Purging Conditions

Column Size : 3 μ m, 2.1 x 150mm
 Eluent : 50mM TEA
 (pH 9.0 or pH 10.0) /
 CH₃OH : 70 / 30, v / v
 Col. Temp : 50°C

—●— pH 9.0
 —■— pH 10.0

Joint for Inlet

- Waters Type

General type


Lineup of 5um and 3um are used for Waters type



- UP Type (For UHPLC)

Lineup of 1.9um and 3um HP are used for UP type

Waters type



UP type

